

SUPPLEMENTARY MATERIAL

Supplementary Material

Table i. Summary of distal femur preparation and implant placement

Operation	Pilot hole, mm	Condensation steps, mm				Implant diameter, mm
		Step 1	Step 2	Step 3	Step 4	
Drilled bone	0.80 (100 rpm)	N/A	N/A	N/A	N/A	0.80
Condensed bone	0.60 (100 rpm)	0.58 to 0.65	0.63 to 0.70	0.68 to 0.75	0.73 to 0.80	0.80

N/A, not applicable.